



**Govt. MEDICAL COLLEGE**  
**THIRUVANANTHAPURAM -695011**

Phone: (0471) 2528055

Fax: (0471) 2443095

E-mail: principalgmct@gmail.com

store.gmct@gmail.com

**PRINCIPAL**

C8/31281/2022/GMCT

Date: 16/11/2022

## QUOTATION NOTICE

Quotation Number	C8/31281/2022/GMCT
Due Date and Time of receipt of Quotations	Date: 29.11.2022 Time: 02.30 PM
Due Date and Time of opening Quotations	Date: 29.11.2022 Time: 03.00 PM
Date up to which the rates are to remain firm of acceptance	31.03.2023
Designation and address of officer to whom the Quotations is to be addressed	The Principal, Govt. Medical College, Thiruvananthapuram - 695 011
<b>Superscriptions:</b> Quotation for the purchase of <b>Multiparameter Monitor</b> for the Department of Obstetrics and Gynecology, SAT Hospital, Govt Medical College, Thiruvananthapuram <b>Specifications: Separate Sheet attached</b>	

Sealed Quotations are invited for the supply of the materials specified. The rates quoted should be for delivery of the articles at the place mentioned below the schedule. The necessary superscription, due to which the rates will have to remain firm for acceptance and the name and address of officer to whom the Quotations is to be sent are noted above. Any Quotation received after the time fixed on the due date is liable to be rejected. The maximum period required for delivery of the articles should also be mentioned. Quotations not accordance with conditions are liable to be rejected.

**Acceptance of the Quotation will subject to the following conditions:**

1. Acceptance of the Quotation constitutes a concluded contract. Nevertheless, the successful tenderer must within a fortnight/a month after the acceptance of his Quotation furnished 5 percentage of the amount of the contract as Security Deposit and execute an agreement at his own cost for the satisfactory fulfillment of the contract, if any.
2. withdrawal from the Quotations after it is accepted or failure to supply within a specified time according to specifications will entail cancellation of the order and purchase being made at the offerers expenses from elsewhere, any loss incurred there by



being payable by the defaulting party. In such an event, the Government also reserves the right to remove the defaulter's name from the list of Government suppliers permanently or for a specified number of years.

(a) C  
taken to  
documents  
with

3. Samples, duly listed, should be forwarded if called for under separate cover and the un approved samples got back as early as possible by the offerers at their own expenses and the Government will in no case be liable for any expenses on account of the value of the samples or their transport charges, etc. In case, the samples are sent by railway, the railway receipt should be sent separately, and not along with the Quotation since Quotation will be opened only on the appointed day and demurrage will have to be paid if the railway parcels are not cleared in time. Quotations for the supply of materials tendered for are forwarded. The approved samples may or may not be returned at the discretion of the undersigned. Samples sent by V P Post or "freight to pay "will not be accepted.

4. No representation for enhancement of Price once accepted will be considered during the currency of the contract.

5. Any attempt on the part of Tenderers or their agents to influence the Officers concerned in their favour by personal canvassing will disqualify the tendered.

6. If any license or permit is required, tenderers must specify in their Quotation and also state the authority to whom application is to be made.

7. The Quotation may be for entire or part supplies. But the tenderers should be prepared to carry out such portion of the supplies included in their Quotation as may be allotted to them.

8. (a) In case, where a successful tenderer, after having made partial supplies fails to fulfill the contracts in full, all or any of the materials not supplied may, at the discretion of the purchasing officer, be purchase by means of another tender/ Quotation or negotiation or from the next higher tenderer who had offered to supply already and the loss, if any, caused to the Government shall there by together with such sums as may be fixed by the Government towards damages be recovered from the defaulting tenderer.

(b) Even in case where no alternate purchase are arranged for the materials not supplied, the proportionate portion of the Security Deposit based on the cost of the materials not supplied at the rate shown in the tender of the defaulter shall be forfeited and balance alone shall be refunded.

(c) Any some of money due and payable to the contractor (Including Security Deposit returnable to him) under this contract may be appropriated by the purchasing officer or Government or any other person authorized by Government and set - off against any claim of the purchasing officer or Government for the payment of a sum of money arising out of or under any other contract made by the contractor with the purchasing officer or Government or any other person authorized by Government.

9. The prices quoted should be inclusive of all taxes, duties, cesses etc., which are or may become payable by the contractor under existing or future laws or rules of the country of origin/supply or delivery during the course of execution of the contract.

also res  
suppl

(a) Ordinarily payments will be made only after the supplies are actually verified and taken to stock but in exceptional cases, payments against satisfactory shipping documents including certificates of insurance will be made up to 90 per cent of the value of the materials at the discretion of the government. Bank charges incurred in connection with payment against documents through bank will be to the account of the contractor. The firms will produce stamped pre receipted invoices in all cases where payments (Advance/Final) for release of railway receipts/shipping document are made through banks. In exceptional cases where the stamped receipts of the firms are not received for the payments (In advance) the unstamped receipts of the bank (ie. counterfoils of pay-in-slips issued by the bank) alone may be accepted as a valid proof for the payment made.

(b) The tenderers shall quote also the parentage of rebate (Discount) offered by them in case the payment is made promptly within fifteen days/within one month of taking delivery of stores.

11. Any sum of money due and payable to the successful tenderer or contractor from Government shall be adjusted against any sum of money due to Government from him under any other contracts.

12. Special conditions, if any, printed on the Quotation sheets of the tenderer or attached with the tenderer will not be applicable to the contracts unless they are expressly accepted in writing by the purchase.

Place: Thiruvananthapuram

Date: 16/11/2022

  
For **PRINCIPAL**



# Technical Specification

## **Equipment : Multiparameter Monitor**

Equipment Name: Multiparameter Monitor

1. Should be able to monitor 6 channel ECG from 5 lead ECG, SPO2, NIBP, Respiration Rate and temperature.
2. Should be portable with carrying handle.
3. Should have touch screen TFT display with at least 12 inches or higher with at least 6 wave forms and numeric display simultaneously.
4. Should have Lithium ion battery with 3 hours battery back up.
5. Should have keys for quick access to main functions.
6. Should have adult, pediatric and neonatal modes.
7. Should provide prominent prioritized audio, visual alarms for high, low heart rate, SpO2, RR, low battery and lethal arrhythmia
8. Recognition with detailed list (more than 15 parameters) in the monitor to set priority and ST Analysis.
9. NIBP can be taken on manual/auto/stat modes.
10. Should provide following accessories:
  - i. Reusable adult 5 lead ECG cable set – 2 nos.
  - ii. Reusable adult and pediatric SPO2 finger probes – 1 each.
  - iii. Adult and pediatric NIBP cuff of different size
  - iv. Rate of Reusable Neonatal SpO2 probe shall be quoted separately which will not be taken for evaluation.
11. Should work on 200-240V AC/50Hz with inbuilt rechargeable battery.

12. Should have safety certificate from a competent authority CE issued by a notified body registered in European Commission / FDA (US) / STQC CB certificate/ STQC S certificate or valid detailed electrical and functional safety test report from ERTL.

Copy of the certificate / test report shall be produced along with the technical bid.

13. Should have display perfusion index

14. Monitor should have networking facility with bidirectional & bed to bed communication.

15. Monitor stand should be supplied along with the machine as per the following specification

- i. Monitor stand -extruded Aluminum, powder coated.
- ii. Load bearing capacity 20 kgs approximately.
- iii. Should supply, install with necessary anchor fastners at the site.

16. Equipment performance should not be affected by electromagnetic interferences radiated or conducted through power lines from another device

PRINCIPAL  
4  
15/11  
2022



*Signature*  
D. Syjamil Jacob  
Prof & HOD  
OBG  
SATH. 3/11/22

**PROFESSOR & HEAD**  
O & G Department  
SAT Hospital, Medical Coll  
Thiruvananthapuram