



Govt. MEDICAL COLLEGE
THIRUVANANTHAPURAM -695011

Phone: (0471) 2528055

Fax: (0471) 2443095

E-mail:principalgmct@gmail.com

PRINCIPAL

C8/24368/2022/GMCT

Date: 19/08/2022

QUOTATION NOTICE

Quotation Number	C8/24368/2022/GMCT
Due Date and Time of receipt of Quotations	Date: 31.08.2022 Time: 02.30 PM
Due Date and Time of opening Quotations	Date: 31.08.2022 Time: 03.00 PM
Date up to which the rates are to remain firm of acceptance	31.03.2023
Designation and address of officer to whom the Quotations is to be addressed	The Principal, Govt. Medical College, Thiruvananthapuram - 695571
Superscriptions: Quotation for the purchase of Ice Lined Refrigerator and Digital Weighing Machine for the Department of Nephrology, Govt. Medical College, Thiruvananthapuram Specifications (Separate sheets attached)	

Sealed Quotations are invited for the supply of the materials specified. The rates quoted should be for delivery of the articles at the place mentioned below the schedule. The necessary superscription, due to which the rates will have to remain firm for acceptance and the name and address of officer to whom the Quotations is to be sent are noted above. Any Quotation received after the time fixed on the due date is liable to be rejected. The maximum period required for delivery of the articles should also be mentioned. Quotations not accordance with conditions are liable to be rejected.

Acceptance of the Quotation will subject to the following conditions:

1. Acceptance of the Quotation constitutes a concluded contract. Nevertheless, the successful tenderer must within a fortnight/a month after the acceptance of his Quotation furnished 5 percentage of the amount of the contract as Security Deposit and execute an agreement at his own cost for the satisfactory fulfillment of the contract, if any.
2. withdrawal from the Quotations after it is accepted or failure to supply within a specified time according to specifications will entail cancellation of the order and purchase being made at the offerers expenses from elsewhere, any loss incurred there by

Technical Specification

Equipment :Ice Lined Refrigerator

1. Temperature Range : 2C to 8C
2. Capacity : 300 L
3. Inner Chamber : Stainless steel 304, 18SWG (1.2mm), tig welded to the middle chamber
4. Middle Chamber : Galvanized steel
5. Outer Chamber : Made of powder coated CRCA sheet, with latch, lock & key
6. Insulation : 75mm thick PUF insulation between the middle and outer chamber
7. Sub doors : Made of polyurethane of 40mm thickness
8. Hot line : Anti condensation using the hot line at the mouth of the cabinet
9. Compressor : Hermetically sealed compressor
10. Refrigerant : Non -CFC, non HCFC, eco friendly (R-134A)
11. Coolant : Brine of ethylene Glycol
12. Combi-condensor : Skin condenser on inner lining of external body+ air cooled type condenser with grooved aluminium fins for effective heat transfer
13. Copper coils : 99.97% pure copper coils
14. Controller : Microprocessor based, with digital display
15. Sensor : Fast response NTC Thermistor sensor
16. Alarm : Audio visual alarm for temp rise, low battery
17. Battery backup during : Inverter circuit for continuous display of power failure temperature and recording (10 to 12 hours approx) with auto recharging, battery ok indication
18. Safety : Automatic Hi/Low voltage cut-off and time delay restart
19. During power failure, temperature of the ILR should be hold for 24 hours - from 2 degree Celsius to 10 degree Celsius

Date: 29/07/2022



Kale
19/8
PRINCIPAL
MEDICAL COLLEGE
THIRUVANANTHAPURAM

4/8/2022

HEAD OF THE DEPARTMENT *2/0*

Dr. NOBLE GRACIOUS S.S.MD,DM
EXECUTIVE DIRECTOR

Kerala State Organ & Tissue Transplant Organisation (K-SOTTO)
1st Floor, Old Hospital Building Quarters,
Near St. Joseph's Hospital Block
Govt. Medical College, Thiruvananthapuram-11
Ph: 0471-2578000, 2572748

Technical Specification

Equipment :Electronic Adult Weighing Scale (Platform Type)


1. Capacity : 160 kg
2. Accuracy : 100 g
3. Platter Size : 350 mm x 300 mm (Tolerance +/- 10%)
4. The scale should be made up of heavy duty. Cast iron structure Platform with powder coated frames.
5. The Electronic Adult Weighing Scale should incorporate following features for user-friendly convenience.
6. Display: LED / LCD : 5 digit with min. height 14 mm.
7. TARE facility with zero function.
8. The Scale should have inbuilt rechargeable battery backup for minimum of 8 hrs.
9. Should operate on mains 220-240Vac, 50 Hz single phase.
10. The Scale should be as per BIS specifications. The scale should have ISI mark.
11. The display stand height shall be 80cm(Tolerance +/- 10%) from the platform.
12. The equipment shall be supplied with valid stamping from Weight and Measures. The stamping shall be done free of cost during the warranty period and the CAMC rates offered shall include the stamping charges.
13. The tenderer shall have valid sales and service license for Weighing Machines from Legal Metrology

Date: 29/07/2022

HEAD OF THE DEPARTMENT



Kale
9/8
PRINCIPAL
MEDICAL COLLEGE
THIRUVANANTHAPURAM



Dr. NOBLE GRACIOUS S.S,MD,DM
EXECUTIVE DIRECTOR
Kerala State Organ & Tissue Transplant Organisation (K-SOTTO)
1st Floor, Old House Surgeon Quarters,
Near Super Speciality Block
Govt. Medical College, Trivandrum-11
Ph: 0471-2528658, 2962748