



സർക്കാർ മെഡിക്കൽ കോളേജ്,  
തിരുവനന്തപുരം, കേരളം - 695011

ഫോൺ: (0471)2528855,2528055  
www.tmc.kerala.gov.in  
principalgmct@gmail.com  
store.gmct@gmail.com

No.C8/2930/2025/GMCT

27-12-2025

**പുനർ ദർഘാസ് പരസ്യം**

തിരുവനന്തപുരം, സർക്കാർ മെഡിക്കൽ കോളേജ്, ട്രാൻസ്ഫ്യൂഷൻ മെഡിസിൻ വിഭാഗത്തിലേക്ക് **Plasma Twawa Bath** വാങ്ങുന്നതിനു നിശ്ചിത ഫോറത്തിൽ മുദ്ര വച്ച ദർഘാസുകൾ ക്ഷണിക്കുന്നു. ദർഘാസ് അടക്കം ചെയ്യുന്ന കവറിനു മുകളിൽ ദർഘാസ് നമ്പർ എഴുതേണ്ടതും, പ്രിൻസിപ്പൽ, സർക്കാർ മെഡിക്കൽ കോളേജ്, തിരുവനന്തപുരം എന്ന വിലാസത്തിൽ അയക്കേണ്ടതുമാണ്. ദർഘാസുകൾ താഴെ ചേർത്തിരിക്കുന്ന തീയതികളിലും, സമയത്തും അപ്പോൾ ഹാജരായിട്ടുള്ള ദർഘാസ് സമർപ്പിച്ചവരുടെയോ അവർ ചുമതലപ്പെടുത്തിയിട്ടുള്ള പ്രതിനിധികളുടെയോ സാന്നിധ്യത്തിൽ തുറക്കുന്നതാണ്. വൈകി കിട്ടുന്ന ദർഘാസുകൾ പരിഗണിക്കുന്നതല്ല.

ദർഘാസ് സമർപ്പിക്കുവാൻ താത്പര്യമുള്ളവർ മെഡിക്കൽ കോളേജ് ഓഫീസിൽ അപേക്ഷിക്കുന്ന പക്ഷം ആവശ്യമായ ദർഘാസ് ഫോറങ്ങൾ താഴെ കൊടുക്കുന്ന പ്രകാരം തുക അടച്ചാൽ ലഭിക്കുന്നതാണ്. ദർഘാസ് ഫോറങ്ങൾ കൈമാറ്റം ചെയ്യാൻ പാടുള്ളതല്ല. ചെക്കുകൾ, പോസ്റ്റേജ് സ്റ്റാമ്പ്, ഡിമാൻഡ് ഡ്രാഫ്റ്റ്, എന്നിവ ദർഘാസ് ഫോറത്തിന്റെ വിലയായി സ്വീകരിക്കുന്നതല്ല. ദർഘാസ് ഫോറം വി.പി.പി. ആയി അയക്കുന്നതല്ല.

അടങ്കൽ തുക : 120,000/- രൂപ

ദർഘാസ് ഫോറത്തിന്റെ വില

(ഒറിജിനൽ - നികുതി സഹിതം) : ₹672/- രൂപ (600/-+72/-)

(ഡ്യൂപ്ലിക്കേറ്റ് - നികുതി സഹിതം): 336 രൂപ

നിരതദ്രവ്യം : ₹1,200/- രൂപ

ഫോറം വിൽക്കപ്പെടുന്ന അവസാന തീയതി, സമയം : 27/01/2026 03.00 PM

ഫോറം സ്വീകരിക്കുന്ന അവസാന തീയതി, സമയം : 28/01/2026 02.30 PM

ഫോറം തുറക്കുന്ന തീയതി,

സമയം: 28/01/2026 03.00 PM

ദർഘാസ് നിരക്ക് പ്രാബല്യത്തിലിരിക്കുന്ന തീയതി : 31-03-2026

ദർഘാസ് നമ്പർ	സാധനങ്ങളുടെ പേര്	യൂണിറ്റ്	അളവ്
14/2025	Plasma Twawa Bath (Specification attached)	എണ്ണം	1

### Technical Specification Plasma Thaw bath

1. Should be able to thaw a minimum of 12 regular Plasma bags.
2. Should be able to thaw plasma bags at present temperature of  $37 \pm 0.2$  C.
3. Should have programmable temp range + 37 C to +56 C.
4. Should have removable stainless - steel tray with individual compartment for holding 12 regular plasma bags. Should have the following Alarms and indicators.
  - a. High/ Low temperature - Audible/Visible
  - b. Timer - Audible/Visible
5. Should have a uniform and optimum thawing mechanism - pumping mechanism by high capacity pump.
6. Should have a system to drain the chamber.
7. Should be supplied with a SS cover with built in handles to keep the unit covered when not in use.
8. Should have a LED display for Set temperature, Actual Temperature.
9. Should have a timer with range of 1 to 240 minutes.
10. The inner chamber should be made of stainless steel 304.
11. External cabinet should be made of 1.2mm HDGI sheet with powder coating.
12. Total weight should not exceed 35 Kg, without water.
13. Should have a PT 100 sealed Sensor dipped directly in the water
14. Should have a display resolution of 0.1 C
15. Should have a microprocessor based digital controlled to control temperature.
16. Should have an external dimension, not greater than 800x500x400mm (LxWxH)
17. Maximum power consumption should be 1100 W
18. Input supply: 230V $\pm$ 10%, 50Hz, Single - Phase AC
19. Should have ISO 9001:2015, ISO 13483 2016 Certificates.
20. Warranty - 3 years

21. Good performance certificate from 3 high volume Government Blood Centres in Kerala to be attached.

22. Demo of the equipment required to ensure the work efficiency of the machine in high volume working conditions.

കൂടുതൽ വിവരങ്ങൾ പ്രവൃത്തി സമയത്ത് ഈ ഓഫീസിൽ അന്വേഷിച്ചാൽ അറിയാവുന്നതാണ്. ടെൻഡർ നടപടികൾ മുന്നറിയിപ്പില്ലാതെ നീട്ടി വയ്ക്കുന്നതിനോ, നിർത്തി വയ്ക്കുന്നതിനോ ഉള്ള അധികാരം പ്രിൻസിപ്പലിൽ നിക്ഷിപ്തമായിരിക്കും.

Signed by Dr Jabbar P K  
Jabbar Puthiyaveetil:Khadar  
#ApprovedByDesignation#  
Date: 27-12-2025 08:36:17

# വിൻഡോ പരസ്യം

സി8/2930/2025/ജി.എം.സി.റ്റി

തീയതി: 27-12-2025

## സർക്കാർ മെഡിക്കൽ കോളേജ്

തിരുവനന്തപുരം

പുനർ ദർഘാസ് നമ്പർ: **14/2025**

തിരുവനന്തപുരം, സർക്കാർ മെഡിക്കൽ കോളേജിലെ ട്രാൻസ്പ്ലാഷ്

മെഡിസിൻ വിഭാഗത്തിലേക്ക് **Plasma Thawa Bath** വാങ്ങുന്നതിന്.

അടങ്കൽ തുക : 1,20,000/- രൂപ

ദർഘാസ് വില : 672/- രൂപ (600+72)

നിരതദ്രവ്യം : 1,200/- രൂപ

ഫോറം വിൽക്കപ്പെടുന്ന അവസാന തീയതി, സമയം : 27-01-2026 03.00 PM

ഫോറം സ്വീകരിക്കുന്ന അവസാന തീയതി, സമയം : 28-01-2026 02.30 PM

ഫോറം തുറക്കുന്ന തീയതി, സമയം : 28-01-2026 03.00 PM

Dr Jabbar P K  
PRINCIPAL

Signed by

Jabbar Puthiyaveettil Khadar

Date: 27-12-2025 08:35:49

**No. GMCTVM/2930/2025-C8**

**27-12-2025**

**Re-Tender No. 14 /2025.**

**RE-TENDER SCHEDULE**

1. The period for which the rates to be in firm - One year.
2. The rates should be inclusive of sales tax and for free delivery at the department concerned, otherwise the tender will be rejected
3. Advance payments conditions will not be considered.
4. Tender forms are neither transferable nor the cost of it will be refundable under any circumstances.
5. Tenders are to be sent with details and illustrated pamphlet (where ever necessary) neatly typed.  
Manuscripts should be avoided as far as possible.
6. Tenders should be accompanied with a formal agreement executed in Rs. 200/- Kerala - non judicial - stamp Paper together with required sum of Earnest Money Deposits.
7. Earnest Money Deposits (one Percentage) of the total cost of the articles quoted for, can be remitted by way of crossed Demand Drafts drawn in favor of the Principal, Medical College, Thiruvananthapuram.
8. Firms will be exempted from furnishing Earnest Money Deposit if they had produced the duly attested copies of relevant certificate along with their tenders.
9. Tenders Unsealed, Unsigned or submitted without superscription as specified are liable to be rejected.
10. Tenders submitted with insufficient Earnest Money Deposit will not be entertained.
11. Belated tenders will not be considered under any circumstance.
12. For Medicines, Consumables, Disposables and Materials, Samples should be submitted along with the Tender. Tenders without samples will not be considered for selection.
13. Tenders should bear the Signature and full Postal address of the Tenderers where ever necessary.
14. The undersigned reserves the right to reject any Tender wholly or partially without any prior information to the Tenders.
15. Specifications of the items are given in the list.
16. Negotiation will be made only with the firm offering lowest price whose specification conforms to the requirements in the Tender Notification.
17. Tender will be opened in the presence of such of the Tenderers or his authorized representative who may present at the time. They should have valid authorization letters
18. More details can be had from this office on working hours.
19. Amount quoted should be in Indian Rupees

PRINCIPAL

Signed by

Jabbar Puthiyaveetil Khadar

Date: 27-12-2025 08:35:19